



E-mail:sales@sunlongbiotech.com

www.sunlongbiotech.com

Shikimate Dehydrogenase (SD) Activity Assay Kit

Note: It is necessary to predict 2-3 large difference samples before the formal determination.

Operation Equipment: Spectrophotometer

Cat No: AK0106 **Size:** 50T/48S

Components:

Extract solution: Liquid 60 mL×1, store at 4°C. Contains insoluble matter, shake well before use.

Reagent I: Liquid 20 mL×1, store at 4C;

Reagent II: Powder×1, store at 4C and protect from light. Add 10 mL distilled water when the solution

will be used;

Reagent III: Powder×2, store at -20°C and protect from light. Add 11 mL distilled water when the solution

will be used.

Product Description:

Shikimic acid pathway is an important metabolic pathway in plants and microorganisms. Shikimate dehydrogenase (EC 1.1.1.25) is the key enzyme to catalyze the fourth step of shikimate synthesis and metabolism.

Shikimate dehydrogenase catalyzes the production of NADPH from shikimate and NADP. The SD activity can be expressed by measuring the increasing rate of absorbance value at 340 nm.

Required but Not Provided:

Ultraviolet spectrophotometer, desk centrifuge, water-bath, transferpettor, 1 mL quartz cuvette, mortar/homogenizer, ice and distilled water.

Protocol

I. Preparation:

- 1. Tissue: according to the ratio of mass (g): extract solution volume (mL): 1:5- 10 to add the extract. It is suggested that add 1 mL of extract solution to 0.1 g of tissue. Homogenate on ice. Centrifuge at 8000 g 4C for 10 min. Take the supernatant on ice for test.
- 2. Cells: first collect bacteria or cells into the centrifuge tube. Discard the supernatant after centrifugation. According to the number of the cells (10⁴): the volume of the extract solution (mL) is 500~1000:1. It is suggested that add 1 mL of extract solution to 5 million of cells. Breaking cells by ultrasonic wave in ice bath (power 300W, ultrasonic 3s, interval 7s, total time 3 min). Centrifuge at 8000 g 4C for 10 min. Take the supernatant on ice for test.
- 3. Liquid: direct detection.

II. Determination procedure:

1. Preheat Spectrophotometer for 30 minutes, adjust wavelength to 340 nm, set the counter to zero with



E-mail:sales@sunlongbiotech.com

www.sunlongbiotech.com

distilled water.

2. Working solution: according to the volume ratio of reagent II: reagent III=7:4:8 to prepare. Mix thoroughly and preheat for 15 min at 25°C.

3. Operation table: add the following reagents to the 1 mL quartz cuvette respectively

Reagent (μL)	Test tube (A _T)	Blank tube (A _B)
Working solution	950	950
Sample	50	
Distilled water		50

Start timing when adding samples. Mix thoroughly. Measure the absorbance value A1 at 340 nm for 20s and A2 at 5min 20s. $\Delta A_B = A2_B - A1_B$. $\Delta A_T = A2_T - A1_T$. $\Delta A = \Delta A_T - \Delta A_B$. Blank tube just needs to test once or twice

III. SD Calculation:

1) Protein concentration:

Unit definition: One unit of enzyme is defined as the amount of enzyme that catalyzes the generation of 1 nmol NADPH per minute every mg tissue protein in the reaction system.

SD (U/mg prot) =
$$\Delta A \div (\epsilon \times d) \times V_{RT} \div (Cpr \times V_{SA}) \div T \times 10^9 = 643 \times \Delta A \div Cpr$$

2) Sample weight:

Unit definition: One unit of enzyme is defined as the amount of enzyme that catalyzes the generation of 1 nmol NADPH per minute every gram tissue weight in the reaction system.

SD (U/g weight) =
$$\Delta A \div (\epsilon \times d) \times V_{RT} \div (W \div V_E \times V_{SA}) \div T \times 10^9 = 643 \times \Delta A \div W$$

3) Cells

Unit definition: One unit of enzyme is defined as the amount of enzyme that catalyzes the generation of 1 nmol NADPH per minute every 10⁴ cells in the reaction system.

SD (U/10⁴ cell) =
$$\Delta A \div (\epsilon \times d) \times V_{RT} \div (cell \ amount \times V_{SA} \div V_E) \div T \times 10^9 = 643 \times \Delta A \div cells$$

4) Liquid volume

Unit definition: One unit of enzyme is defined as the amount of enzyme that catalyzes the generation of 1 nmol NADPH per minute every mL liquid in the reaction system.

SD (U/mL)=
$$\Delta A \div (\epsilon \times d) \times V_{RT} \div V_{SA} \div T \times 10^9 = 643 \times \Delta A$$

ε: NADPH molar extinction coefficient, 6.22×10³ L/mol/cm;

d: Light path of cuvette, 1 cm;

V_{RT}: Total reaction volume, 0.001 L;

V_{SA}: Sample volume, 0.05 mL;

V_E: Extract solution volume, 1 mL;

T: Reaction time, 5 min;

Cpr: Protein concentration, mg/mL;

W: Sample weighr, g;

10⁹: Unit conversion coefficient, 1 mol=10⁹ nmol.

Cell amount: 10⁴ cells as a unit.

SunLong Biotech Co.,LTD





E-mail:sales@sunlongbiotech.com

www.sunlongbiotech.com

Note:

- 1. The supernatant of sample extraction should be placed on ice for test, and it is recommended to complete the test within 2 hours after sample extraction.
- 2. The protein concentration of the sample needs to be determined by yourself. Because the extract contains a certain concentration of protein (about 1 mg/mL), the protein concentration of the extract solution should be deducted when determining the protein concentration of the sample.
- 3. When ΔA is greater than 1, it is recommended to dilute the sample before determination. When ΔA is less than 0.01, the reaction time (10 min or 15 min) can be prolonged to determine.
- 4. Blank tube is the test tube for testing the quality of each reagent component. Under the normal conditions, the change of blank tube shall not exceed 0.01.

Experimental Examples:

1. Take 0. 1g of radish leaves, add 1mL extract, and process the sample. Take the supernatant and follow the measurement steps. The calculation is calculated as $\Delta At = A2t - A1t = 1.441 - 1.201 = 0.24$, $\Delta Ab = A2b - A1b = 0.24$, calculated according to sample quality:

SD activity (U/g mass) = $643 \times \Delta A \div W = 1543.2 \text{ U/g} \text{ mass}$.

Related Products:

AK0246/AK0245 Isocitrate Lyase (ICL) Activity Assay Kit AK0149/AK0148 Acetokinase (ACK) Activity Assay Kit