

Water Mercury Ion (Hg^{2+}) Content Assay Kit

Note: Take two or three different samples for prediction before test.

Operation Equipment: Spectrophotometer/ Microplate Reader

Cat Number: AK0176

Size: 100T/96S

Components:

Reagent I: Powder $\times 1$. Storage at 4°C . Dissolve with 2 mL of distilled water before use.

Reagent II : Liquid 5 mL $\times 1$. Storage at 4°C .

Reagent III: Liquid 10 mL $\times 1$. Storage at 4°C .

Reagent IV: Powder $\times 1$. Storage at 4°C . Dissolve with 5 mL of distilled water before use.

Reagent V: Powder $\times 1$. Storage at 4°C and protect from light. Dissolve with 50 mL of chloroform (**self-provided**) before use.

Reagent VI: Liquid 20 mL $\times 1$. Storage at 4°C .

Standard Solution: Liquid 1 mL $\times 1$, 4000 nmol/mL Hg^{2+} standard solution. Storage at room temperature. Add distilled water dilute 400 times to form a standard solution of 10 nmol/mL before use.

Product Description:

Hg^{2+} is an important toxic heavy metal ion in water, which easily absorbed and accumulated by organisms and can be further transmitted through the food chain, causing damage. Minamata disease is a kind of typical mercury poisoning.

After digestion, Hg^{2+} can form one orange complex with Dithizone in acid environment, which can be dissolved in chloroform and has a maximum absorption peak at 490 nm. In this kit, the content of Hg^{2+} is quantified by measuring the color development at 490 nm.

Reagents and Equipment Required but Not Provided.

Spectrophotometer/microplate reader, centrifuge, adjusted transferpettor, micro glass cuvette/96 well flat-bottom plate, concentrated sulfuric acid (H_2SO_4), concentrated nitric acid (HNO_3), distilled water.

Operation steps:

I . Sample processing

Add 7 mL of concentrated nitric acid immediately after every 1000 mL of water sample is collected. Adjust the $\text{pH} \leq 1$. If the water sample cannot be measured immediately after sampling, add 4 mL or more of Reagent II to each liter of sample to make it lasting pale red.

II . Determination

1. Preheat the spectrophotometer/microplate reader for 30 minutes, adjust wavelength to 490 nm, set zero with chloroform.

2. Add reagents with the following list to 1.5 mL EP tubes:

Reagent Name (μL)	Test tube (A _T)	Standard tube (A _S)	Blank tube (A _B)
Water sample	400	-	-
Standard solution	-	400	-
Distilled water	-	-	400
Concentrated sulfuric acid	16	16	16
Concentrated nitric acid	4	4	4
Reagent I	13	13	13
Reagent II	24	24	24
Seal with parafilm, mix thoroughly and shock 2 minutes. Digest in 95°C water bath for 2 hours, then cool to about 40°C .			
Reagent III	80	80	80
Shake until the solution in the EP tube is clear and transparent. Open the lid and leave for 10 minutes. Shake several times during standing to allow the gas escape.			
Reagent IV	32	32	32
Reagent V	400	400	400
Fully shake for 2 minutes after capping, stand for 10 minutes. Suck the organic phase in the lower layer into 1.5 mL EP tube.			
Reagent VI	160	160	160
Fully concussion to make the organic phase without green. After standing and delaminating, take 200μL of the organic phase and measure the absorbance at 490 nm. Recorded as A _T , A _S , A _B . ΔA _T =A _T -A _B , ΔA _S =A _S -A _B .			

Calculations

$$\text{Hg}^{2+} (\text{nmol/mL}) = C_s \times \Delta A_T \div \Delta A_S = 10 \times \Delta A_T \div \Delta A_S$$

Cs: Hg²⁺

standard solution (10 nmol/mL).

Note:

- 1000 μg/L copper ion, 20 μg/L silver ion, 10 μg/L gold ion, 5 μg/L platinum ion in water sample without interference.
- Pay attention to safety during measurement, wear masks and gloves to avoid inhalation or exposure to toxic and dangerous reagents.
- When the absorbance is greater than 0.6, please dilute the serum to appropriate concentration with distilled water.
- Water with less suspended matter and/or organic matter can shorten the heating time to 1 hour, and clean water without suspended matter can shorten the heating time to 30 minutes.
- if the upper solution of sample tube becomes transparent during digestion, Reagent II can be added

appropriately to keep the sample tube pink or black purple.

6. if the added Reagent III is not enough to make the sample tube clear, the amount of Reagent III can be added to make the sample tube clear.

7. if the lower organic phase still appears green after Reagent VI is added, the amount of Reagent VI can be increased to make the organic phase transition of the lower layer shallow.

8. polypropylene/polystyrene is easy to dissolve in organic reagents such as chloroform.

Related Products:

AK0175/AK0174 Water Chromium(Cr^{6+}) Content Assay Kit

AK0376/AK0375 Total Phosphorus Content Assay Kit

AK0078/AK0077 Tissue Iron Content Assay Kit

AK0074/AK0073 Blood Ammonia Content Assay Kit