

SUNLONG

E-mail:sales@sunlongbiotech.com

www.sunlongbiotech.com

Ca⁺⁺Mg⁺⁺-ATPase Activity Assay Kit

Note: Take two or three different samples for prediction before test.

Operation Equipment: Spectrophotometer

Cat No: AK0502 **Size:** 50T/24S

Components:

Reagent I: Liquid 30 mL×1. Storage at 4°C. Reagent II: Liquid 4 mL×1. Storage at 4°C.

Reagent III: Powder \times 2. Storage at -20°C . Dissolve thoroughly with 1 mL of distilled water before use. The rest reagent can be kept at -20°C for one week.

Reagent IV: Liquid 2 mL×1. Storage at 4°C.

Reagent V: Powder×1. Storage at 4°C. Dissolve thoroughly with 3 mL of distilled water before use.

Reagent VI: Powder $\times 1$. Storage at 4°C . Dissolve thoroughly with 15 mL of distilled water before use, can be kept at 4°C for one week.

Reagent VII: Powder \times 1. Storage at 4°C . Dissolve thoroughly with 15 mL of distilled water before use, can be kept at 4°C for one week.

Reagent VIII: Liquid 15 mL×1. Storage at RT.

Standard solution: Liquid 1 mL×1. 10 µmol/mL standard phosphorus liquid, storage at 4°C.

0.5 μ mol/mL standard phosphorus working solution: Dilute the 10 μ mol/mL standard 20 times with distilled water to 0.5 μ mol/mL standard. For example: add 1.9 mL of distilled water to 0.1 mL of standard, mix thoroughly.

Phosphorus fixing reagent:

Prepare reagents for determining phosphorus content: make solution as the volume ratio of H_2O : Reagent VII: Reagent VIII =2:1:1:1, which should be light yellow. It shows lose efficacy if color is changed, phosphorus pollution if color is change to blue. Prepare the reagent when it will be use.

Note: It is better to use new beakers, glass rods and glass pipettes or disposable plastic ware when making reagent to avoid phosphorus pollution.

Product Description:

Ca⁺⁺Mg⁺⁺-ATPase is widely distributed in plants, animals, microorganisms and cells, which catalyzes the hydrolysis of ATP to form ADP and inorganic phosphorus.

Ca⁺⁺Mg⁺⁺-ATPase decomposes ATP to produce ADP and inorganic phosphorus. The activity of ATPase can be detected by measuring the amount of inorganic phosphorus.

Reagents and Equipment Required but Not Provided:

Spectrophotometer, desk centrifuge, adjustable pipette, water bath, 1 mL glass cuvette, mortar/homogenizer, ice and distilled water.



E-mail:sales@sunlongbiotech.com

www.sunlongbiotech.com

Procedure:

I. Sample preparation:

1. Bacteria or cells:

Collecting bacteria or cells into a centrifuge tube, centrifugation and discard supernatant. Suggest add 1mL of Reagent I to 5 million of bacteria or cells. Use ultrasonic to splitting bacteria and cells (placed on ice, ultrasonic power 20%, working time 3 seconds, interval 10 seconds, repeat for 30 times). Centrifuge at 8000 ×g for 10 minutes at 4°C and take the supernatant on ice before testing.

2. Tissue:

Add 1 mL of Reagent I into 0.1 g of tissue, fully grinding on ice. Centrifuge at 8000 ×g for 10 minutes at 4°C and take the supernatant on ice before testing.

3. Serum: Detect directly.

II. Determination:

- 1. Preheat spectrophotometer for 30 minutes, adjust the wavelength to 660 nm, set the counter to zero with distilled water.
- 2. Add the following reagents to EP tube:

Reagent (μL)	Control tube (C) Test tube (T)				
Reagent I	130	90			
Reagent II	80	80			
Reagent III	40	40			
Reagent IV		40			
Sample		200			
Mix thoroughly, then place the reaction solution in a 37°C (mammal) or 25°C (other species) water					
bath for 10 minutes					
Reagent V	50	50			
Sample	200				
Mix thoroughly, centrifuge at 4000 ×g for 10 minutes at room temperature, take the supernatant.					

3. Determination of phosphorus content, add the following reagents to 1.5 mL EP tube:

Reagent (µL)	Blank tube (B)	Standard tube (S)	Control tube (C)	Test tube (T)
0.5 μmol/mL		100		
standard		100		
phosphorus liquid				
Supernatant			100	100
Distilled water	100			
Reagents for	1000	1000	1000	1000
determining				
phosphorus content				

Mix thoroughly, then place the mix solution in a 40°Cwater bath for 10 minutes. Cooling to room temperature and detect the absorbance at 660 nm. The blank tube and standard tube just need one or two tubes.

III. Calculation:



E-mail:sales@sunlongbiotech.com

www.sunlongbiotech.com

1. Serum:

Unit definition: One unit of enzyme activity is defined as the amount of enzyme catalyzes the decompose of ATP to produce 1 μ mol of inorganic phosphorus per hour every milliliter of serum.

$$Ca^{++}Mg^{++}$$
-ATPase (U/mL) = $Cs\times[\Delta A(T)-\Delta A(C)]\div[\Delta A(S)-\Delta A(B)]\times Vrv\div s\div T$

$$=7.5\times[\Delta A(T)-\Delta A(C)]\div[\Delta A(S)-\Delta A(B)]$$

2. Tissue, bacteria or cells

(1) Protein concentration:

Unit definition: One unit of enzyme activity is defined as the amount of enzyme catalyzes the decompose of ATP to produce 1 µmol of inorganic phosphorus per hour every milligram of tissue protein.

$$Ca^{++}Mg^{++}$$
-ATPase (U/mg prot)= $Cs\times[\Delta A(T)-\Delta A(C)]\div[\Delta A(S)-\Delta A(B)]\times Vrv\div(Vs\times Cpr)\div T$

=7.5×
$$[\Delta A(T)-\Delta A(C)]$$
÷ $[\Delta A(S)-\Delta A(B)]$ ÷Cpr

(2) Sample weight:

Unit definition: One unit of enzyme activity is defined as the amount of enzyme catalyzes the decompose of ATP to produce 1 µmol of inorganic phosphorus per hour every milligram of tissue.

$$Ca^{++}Mg^{++}$$
-ATPase (U/g weight)= $Cs\times[\Delta A(T)-\Delta A(C)]\div[\Delta A(S)-\Delta A(B)]\times Vrv\div(Vs\div V1\times W)\div T$

=7.5×[
$$\Delta A(T)$$
- $\Delta A(C)$]÷[$\Delta A(S)$ - $\Delta A(B)$]÷W

(3) bacteria or cells

Unit definition: One unit of enzyme activity is defined as the amount of enzyme catalyzes the decompose of ATP to produce 1 µmol of inorganic phosphorus per hour every 10000 cells or bacteria.

$$Ca^{++}Mg^{++}-ATPase (U/10^{4}cell) = Cs \times [\Delta A(T)-\Delta A(C)] \div [\Delta A(S)-\Delta A(B)] \times Vrv \div (Vs \div V1 \times 500) \div T$$

=0.015×[
$$\Delta A(T)$$
- $\Delta A(C)$]÷[$\Delta A(S)$ - $\Delta A(B)$]

Cs: Concentrate of standard tube, 0.5 µmol/mL;

Vrv: Total reaction volume, 0.5 mL;

Vs: Sample volume, 0.2 mL;

Cpr: Sample protein concentration (mg/mL);

T: Reaction time (min), 1/6 hour;

W: Sample weight(g);

VI: Volume of reagent I, 1 mL;

500: The amount of bacteria or cell, 5 million.

Note:

- 1. This kit can detect 24 tubes of Ca⁺⁺Mg⁺⁺-ATPase samples in 50 tubes for each sample need one tube as control.
- 2. This method has the characteristics of trace, sensitive and rapid. The test tubes used for determination are phosphate-free strictly. Avoiding phosphorus pollution is the key to the success of detection.

Experimental example:

1. Take 0. 1g of pancreas and add 1 mL of Reagent I for ice bath homogenization. After centrifugation at 4°C for 10 min, the supernatant is put on the ice and operated according to the determination steps.

$$\Delta A_T =$$





Tel: 0086-571-56623320 Fax:0086-571-56623318

E-mail:sales@sunlongbiotech.com

www.sunlongbiotech.com

0.916-0.389=0.527, $\Delta A_s = 0.398-0.004=0.394$

Ca++Mg++- ATPase activity (U/g mass) = $7.5 \times \Delta A_T \div \Delta A_S \div W = 100.32$ U/g mass.

2. Take 0. 1g of willow and add 1 mL of Reagent I for ice bath homogenization. After centrifugation at 4°C for 10 min, the supernatant is put on ice and operated according to the determination steps. The ΔA_T =0. 137-0. 124=0.013, and the ΔA_S = 0.398-0.004=0.394

 $Ca + + Mg + + - ATPase activity (U/g mass) = 7.5 \times \Delta A_T \div \Delta A_S \div W = 2.47 U/g mass.$

Recent Product Citations:

[1] Yupu Jing, Hongli An, Shanjing Zhang, et al. Protein kinase C mediates juvenile hormone-dependent phosphorylation of Na⁺. Journal of Biological Chemistry. November 2018;(IF4. 106)

References:

[1] Datiles M J, Johnson E A, McCarty R E. Inhibition of the ATPase activity of the catalytic portion of ATP synthases by cationic amphiphiles[J]. Biochimica et Biophysica Acta (BBA)-Bioenergetics, 2008, 1777(4): 362-368.

Related Products:

AK0602/AK0601 Na+K+——ATPase Activity Assay Kit

AK0309/AK0561 ATP Activity Assay Kit