

Plant Ammoniacal Nitrogen Assay Kit

Operation Equipment: Spectrophotometer

Note: Take two or three different samples for prediction before test.

Catalog Number: AK0426

Size:50T/48S

Components:

Extract solution: 60 mL×1 bottle, storage at 4°C .

Reagent I: powder×1 bottle, storage at 4°C, dissolve with Reagent III thoroughly before use, use up in 10 days.

Reagent II: powder×2 bottle, storage at 4°C and protect from light, dissolve with 2 mL distilled water before use.

Reagent III: 45 mL×1 bottle, storage at 4°C .

Standard: powder×1 bottle, 10 mg cysteine, storage at 4°C and protect from light. Add 1.157 mL Extract solution to make 1000 µg/mL nitrogen standard solution.

Product Description:

Nitrogen is an essential element of living organisms. Ammonium nitrogen enters plant cells and forms amino acids or amides. Ammonia nitrogen content in plant tissues can reflect the degree of plant stress.

Alpha - amino acid can react with hydrated indene triketone to forms blue-purple compound which has characteristic at 570 nm. The amino acid content was calculated by measuring absorbance at 570 nm.

Reagents and Equipments Required but Not Provided:

Spectrophotometer, water bath, centrifuge, transferpettor, 1 mL glass cuvette, mortar/homogenizer, anhydrous ethanol, ice and distilled water.

Procedure:

I. Sample preparation:

Add 1 mL Extract solution into 0.1 g tissue, homogenate at RT, 12000 g 25°C centrifuge for 10 min, take the supernatant on ice is used for test

II. Determination procedure:

1. Preheat spectrophotometer for 30 min, adjust the wavelength to 570 nm, set the counter to zero with distilled water.
2. Dilute 1000 $\mu\text{g/mL}$ nitrogen standard solution with Extract solution to 200 $\mu\text{g/mL}$, 100 $\mu\text{g/mL}$, 50 $\mu\text{g/mL}$, 25 $\mu\text{g/mL}$, 12.5 $\mu\text{g/mL}$ for use.
3. Operational table

Reagent name (μL)	Test tube A_T	Standard tube A_S	Blank tube A_B
Sample	50	-	-
Standard solution	-	50	-
distilled water	-	-	50
Reagent I	500	500	500
Anhydrous ethanol	500	500	500
Reagent II	50	50	50

Mix thoroughly, cover cup with sealing film tightly, keep in boiling water for 10 min, reverse the EP tube some times after cooling, 8000 rpm centrifuge for 5 min, detect absorbance at 570 nm. Detect within 30 min, calculate $\Delta A(\text{standard}) = \Delta A(S) = A_S - A_B$, $\Delta A(\text{test}) = \Delta A(T) = A_T - A_B$.

III. Calculation:

1. Make standard curve:

Nitrogen standard liquid as the abscissa, $\Delta A(S)$ as ordinate, establish the standard curve, get formula $y=kx+b$, take $\Delta A(\text{test})$ to formula, get x ($\mu\text{g/mL}$).

2. Calculation of $\text{NH}_3\text{-N}$ content

A. Sample weight:

$$\text{NH}_3\text{-N} (\mu\text{g/g FW}) = x \times V_e \div W = x \div W$$

B. Protein concentration:

$$\text{NH}_3\text{-N} (\mu\text{g/mg prot}) = x \times V_e \div (C_{pr} \times V_e) = x \div C_{pr}$$

C_{pr} : Protein concentration (mg/mL);

W : Sample weight (g);

Ve: Extract solution volume, 1 mL;

Note:

In order to ensure the accuracy of the experimental results, we need to take 1-2 samples for pre-experiment. If the absorbance is too high (higher than 1), dilute the extract and then determine.

Technical Specifications:

Minimum Detection Limit: 5.5847 $\mu\text{mol/mL}$

Linear Range : 25-300 $\mu\text{mol/mL}$

Recent Product citations:

[1] Fuyuan Zhu, Moxian Chen, Wailung Chan, et al. SWATH-MS quantitative proteomic investigation of nitrogen starvation in Arabidopsis reveals new aspects of plant nitrogen stress responses. Journal of Proteomics. September 2018; (IF3.537)

Related products :

AK0301/AK0300 Nitrate Reductase(NR) Activity Assay Kit

AK0436/AK0435 Glutaminase (GLS) Assay Kit

AK0434/AK0433 Glutamate dehydrogenase (GDH) Activity Assay Kit