



SunLong Biotech Co.,LTD

Tel: 0086-571-56623320 Fax:0086-571-56623318

E-mail:sales@sunlongbiotech.com

www.sunlongbiotech.com

---

# **Human Papillomavirus 16 Major Capsid Protein L1 Antibody (HPV16 L1 protein Ab)ELISA Kit**

**96 Tests**

**Catalogue Number:SLY4385Hu**

**Store all reagents at 2-8℃**

**Validity Period: six months**

**For samples:**

**In serum, plasma, culture media or any biological fluid.**

**FOR RESEARCH USE ONLY !**

**NOT FOR THERAPEUTIC OR DIAGNOSTIC APPLICATIONS !**

**PLEASE READ THROUGH ENTIRE PROCEDURE BEFORE BEGINNING !**

# Human Papillomavirus 16 Major Capsid Protein L1 Antibody (HPV16 L1 protein Ab)ELISA Kit

***FOR RESEARCH USE ONLY***

---

---

## **Drug Names**

Generic Name: **Human Papillomavirus 16 Major Capsid Protein L1 Antibody (HPV16 L1 protein Ab)ELISA Kit**

## **Purpose**

Our Human Papillomavirus 16 Major Capsid Protein L1 Antibody (HPV16 L1 protein Ab) ELISA Kit is to for the qualitative determination of HPV16 L1 protein Ab in Human serum, plasma, culture media or any biological fluid.

## **Intended use**

This HPV16 L1 protein Ab ELISA kit is intended Laboratory for Research use only and is not for use in diagnostic or therapeutic procedures. The Stop Solution changes the color from blue to yellow and the intensity of the color is measured at 450 nm using a spectrophotometer. In order to measure the concentration of HPV16 L1 protein Ab in the sample, this HPV16 L1 protein Ab ELISA Kit includes a set of calibration standards. The calibration standards are assayed at the same time as the samples and allow the operator to produce a cutoff value. The existence or not of HPV16 L1 protein Ab in the samples is then determined by comparing the O.D. of the samples to the CUT OFF.

## **Sample collection and storages**

**Serum** - Use a serum separator tube and allow samples to clot for 30 minutes before centrifugation for 10 minutes at approximately 3000×g. Remove serum and assay immediately or aliquot and store samples at -20°C or -80°C. Avoid repeated freeze-thaw cycles.

**Plasma** - Collect plasma using EDTA or heparin as an anticoagulant. Centrifuge samples for 30 minutes at 3000×g at 2-8°C within 30 minutes of collection. Store samples at -20°C or -80°C. Avoid repeated freeze-thaw cycles.

**Cell culture supernates and other biological fluids** - Remove particulates by centrifugation and assay immediately or aliquot and store samples at -20°C or -80°C. Avoid repeated freeze-thaw cycles.



**Note:** The samples should be centrifuged adequately and no hemolysis or granule was allowed.

## Materials required but not supplied

1. Standard microplate reader(450nm)
2. Precision pipettes and Disposable pipette tips.
3. 37 °C incubator

## Precautions

1. Do not substitute reagents from one kit to another. Standard, conjugate and microplates are matched for optimal performance. Use only the reagents supplied by manufacturer.
2. Do not remove microplate from the storage bag until needed. Unused strips should be stored at 2-8°C in their pouch with the desiccant provided.
3. Mix all reagents before using.

Remove all kit reagents from refrigerator and allow them to reach room temperature ( 20-25°C)

## Materials supplied

Name	96 determinations	48 determinations
Microelisa stripplate	12*8strips	12*4strips
Negative control	0.5ml	0.5ml
Positive control	0.5ml	0.5ml
HRP-Conjugate reagent	10.0ml	5.0ml
20X Wash solution	25ml	15ml
Sample Diluent	6.0ml	3.0ml
Chromogen Solution A	6.0ml	3.0ml
Chromogen Solution B	6.0ml	3.0ml
Stop Solution	6.0ml	3.0ml
Closure plate membrane	2	2
User manual	1	1
Sealed bags	1	1

## Reagent preparation

20×wash solution:Dilute with Distilled or deionized water 1:20.

## **Assay procedure**

1. Prepare all reagents before starting assay procedure. It is recommended that all Standards and Samples be added in duplicate to the Microelisa Stripplate.
2. Separately add Positive control and Negative control 50 $\mu$ l to the Positive and Negative well; Add testing sample 10 $\mu$ l then add Sample Diluent 40 $\mu$ l to testing sample well.
3. Add 100 $\mu$ l of HRP-conjugate reagent to each well, cover with an adhesive strip and incubate for 60 minutes at 37°C.
4. Aspirate each well and wash, repeating the process four times for a total of five washes. Wash by filling each well with Wash Solution (400 $\mu$ l) using a squirt bottle, manifold dispenser or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Solution by aspirating or decanting. Invert the plate and blot it against clean paper towels.
5. Add chromogen solution A 50 $\mu$ l and chromogen solution B 50 $\mu$ l to each well. Gently mix and incubate for 15 minutes at 37°C. **Protect from light.**
6. Add 50 $\mu$ l Stop Solution to each well. The color in the wells should change from blue to yellow. If the color in the wells is green or the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
7. Read the Optical Density (O.D.) at 450 nm using a microtiter plate reader within 15 minutes.

## **Determine the result**

1. Test validity: the average of Positive control well  $\geq 1.00$ ; the average of Negative control well  $\leq 0.15$ .
2. Calculate Critical (CUT OFF): Critical = the average of Negative control well + 0.15.

Negative Result: sample OD < Calculate Critical (CUT OFF) is Negative.

Positive Result: sample OD  $\geq$  Calculate Critical (CUT OFF) is Positive.

## **Storage and validity**

Storage: 2-8°C

Validity: 6 months.